

Comparative Matrix for Extraoral Range

Features & Specifications		KODAK 8000	KODAK 8000C	KODAK 9000	KODAK 9000 3D	KODAK 9000C	KODAK 9000C 3D	KODAK 9500
Panoramic		✓	✓	✓	✓	✓	✓	✓ ***
Cephalometric		Upgradeable	✓	Upgradeable	Upgradeable	✓	✓	✓
3D imaging		✗	✗	Upgradeable	✓	Upgradeable	✓	✓
3D field-of-view size		N/A	N/A	N/A	37 x 50 mm	N/A	37 x 50 mm	90 x 150 mm 180 x 200 mm
Time to process*		Pan: 14 sec.	Pan: 14 sec. Ceph 1 to 3 sec.	Pan: 14 sec. Ceph 1 to 3 sec.	Pan: 14 sec. 3D: 90 sec. to 2 min.	Pan: 14 sec. Ceph: 1 to 3 sec.	Pan: 14 sec. Ceph: 1 to 3 sec. 3D: 90 sec. to 2 min.	1:45 min. 2:15 min.
Footprint/size**	mm	880 (W) x 1180.5 (D)	2250 (W) x 1261 (D)	1158 (W) x 1595 (D)	1158 (W) x 1595 (D)	2150 (W) x 1595 (D)	2150 (W) x 1595 (D)	1416 (W) x 1725 (D)
	inch	34" (W) x 46" (D)	88.59" (W) x 49.63" (D)	45.5" (W) x 62.75" (D)	45.5" (W) x 62.75" (D)	84.63" (W) x 62.75" (D)	84.63" (W) x 62.75" (D)	55.74" (W) x 67.9" (D)

* Acquisition time + reconstruction time. ** Dimensions of the unit. Required operational space could be different. *** Simulated panoramic and cephalometric.

Applications****	KODAK 8000	KODAK 8000C	KODAK 9000	KODAK 9000 3D	KODAK 9000C	KODAK 9000C 3D	KODAK 9500
Routine panoramic examination	●●●	●●●	●●●	●●●	●●●	●●●	●●
Routine cephalometric evaluation		●●●			●●●	●●●	●●●
Implantology	●●	●●	●	●●●	●	●●●	●●●
Endodontic				●●●			
Orthodontic		●●●		●	●●●	●●●	●●
Periodontic	●	●	●	●●●			●●●
Oral Surgery	●	●	●	●●●	●	●●●	●●●
Prosthetic	●	●	●	●●			●●●
Restorative	●	●	●	●●	●	●●	●●●

● Good ●● Better ●●● Best

**** Please note bullets are simply a recommendation based on feedback from our thought leaders and users. One bullet indicates it meets the basic requirements for the application. Three bullets indicates it is the better choice.